RETURN TO PRODUCT BULLETIN TABLE OF CONTENTS
RETURN TO CD-ROM TABLE OF CONTENTS

NORTH AMERICAN DRÄGER

PRODUCT BULLETIN

Product(s):	Narkomed GS
Bulletin Number:	68
Release Date:	10/24/97
Pages:	1

Subject: Narkomed GS Software version 1.1 (Alarm Standby Indication)

The software revision changes the condition in which the "off" indicators are illuminated. When the ventilator is off and the "Alarms Off" keypads are pressed, the "off" lights in the Breathing Volume and Breathing Pressure channels are illuminated. When the ventilator is off and the "All Standby" keypad is pressed, the "on" and "off" indicators (LEDs) are not illuminated (*this is the revision*). There is no change to the system when the ventilator is on -- when the ventilator is on the pressure alarm cannot be turned off or placed in standby.

If you have any questions regarding this bulletin, please call Greg Sutherland at ext. 2311.